

# ASDJ

AINSHAMS DENTAL  
JOURNAL

Print ISSN 1110-7642

Online ISSN 2735-5039

AIN SHAMS DENTAL JOURNAL

Official Publication of Ain Shams Dental School

September 2020 • Vol. XXIII

## Shaping Ability of Different Rotary Nickel Titanium Systems (An In Vitro Study)

*Asmaa AbuBakr Hussien<sup>1</sup>, Abeer Abdel-Hakeem El-Gendy<sup>2</sup>, Tariq Yehia Abdelrahman<sup>3</sup>*

### Abstract

#### Aim

The aim of the study was to evaluate the shaping ability of three rotary nickel titanium systems Revo-S, M-Pro and 2Shape regarding the change of canal curve.

#### Materials and methods

Thirty human mandibular molars were selected for the study, with moderate angle of curvature between (25°-35°). Samples were randomly distributed into three equal groups each of ten according to system used for mesiobuccal canal preparation. Cone beam computed tomography was used to evaluate angle of canal curvature pre and post instrumentation using Schneider method.

#### Results

There was a statistically significant difference between different groups in percentage of reduction of canal curve. The highest values of percentage of reduction was found in Revo-S group.

#### Conclusion

Revo -S root canal rotary system caused the highest percent of reduction in canal curve among all groups.

#### Keywords

Shaping ability, M-Pro, 2Shape, Revo-S

1. Teaching Assistant Faculty of Dentistry, British University in Egypt
2. Professor of Endodontic Faculty of Dentistry, Ain Shams University
3. Lecturer of Endodontics Faculty of Dentistry, Ain Shams University

## Introduction

The main objective of root canal treatment is bacterial eradication through mechanical preparation and canal disinfection, while Preserving original canal anatomy and configuration.

Although Nickel titanium (Ni-Ti) rotary file systems greatly improved root canal preparation efficiency (1), canal preparation and instrumentation is still challenging (2).

Recently introduced, 2Shape (2S) Nickel titanium rotary file is made of alloy called T-wire. 2Shape has a sequence with two instruments: TS1 (#25, 0.04) and TS2 (#25, 0.06) (3).

M-Pro nickel titanium rotary file made of a controlled memory alloy with sequence of three instruments. orifice opener (#18,0.09), two shapers (#20,0.04) and (#25,0.06) (4). Revo-S is made of a conventional nickel titanium alloy.

The aim of this was to evaluate the shaping ability of three rotary nickel titanium systems Revo-S, M-Pro and 2Shape regarding the change of canal curve.

## Materials and methods

Thirty human permanent mandibular molars with mature apices and having moderate curvature ( $25^{\circ}$ - $35^{\circ}$ ) were selected for this study (n=30).

Radiographs were taken to determine degree of curvature according to Scheinder's method (5). Access cavities were prepared and the root canals were negotiated using #10 K-file (Mani, Inc, Tochigi, Japan) The working length of the mesiobuccal canal was measured by inserting K file #10 until it is visible from the apex then subtracting one millimeter from tooth length (6)

The teeth were randomly divided into three equal groups (n=10)

Group (A) Revo-S (n=10)

Group (B) M-Pro (n=10)

Group (C) 2Shape (n=10)

Instrumentation was done according to manufacturer's instruction.

## Cone-beam computed tomography analysis

Pre and post mechanical preparation canal scanning was done using cone beam computed tomography (CBCT) (7) The pre and post -operative angles of curvature were measured according to Schneider technique on CBCT images.

Schneider's method is done by drawing a line parallel to the long axis of the canal in the coronal third (line A). A second line is then drawn from the apical foramen to intersect the point where the canal starts to deviate from the long axis (line B). The angle between them is the Schneider's angle(5) (figure 1).




Figure (1) Schneider method

## Evaluation of canal curvature

Percent of reduction of canal curvature was calculated according to following equation

$$\frac{\text{Pre instrumentation measurement} - \text{post instrumentation measurement}}{\text{Pre instrumentation measurement}} \times 100$$

The greater the percentage the more change happened


## Results

Mean and standard deviation (SD) values of percentage of reduction of canal curvature were presented in table (1) and figure (2)

Results showed that there was a statistically significant difference between different groups in percentage of reduction of canal curve. The highest values of percentage of reduction was found in Revo-S group followed by 2Shape, while the lowest value was for M-Pro group.

**Table (1):** Mean ± standard deviation (SD) of percentage of reduction of canal curvature

Percentage of reduction of canal curve (mean±SD)			p-value
Revo-s	M-pro	2shape	
15.22±2.49 <sup>A</sup>	8.10±1.27 <sup>C</sup>	11.82±3.17 <sup>B</sup>	<0.001*



**Figure (2):** Bar chart showing average percentage of reduction of canal curvature

**Discussion**

the main technical goal during root canal preparation is to develop a continuously tapered preparation from orifice to root canal apex, while maintaining original root canal anatomy and configuration(5) Root canal preparation can be done by using manual or rotary instrumentation. Both instrumentation techniques used in root canal preparation may lead to procedural errors(8). However, several studies showed that Ni-Ti rotary instruments maintained original canal shape better than stainless steel files(9).

With their advantage of simulating the clinical situation extracted human teeth was used rather than stimulated canal (10). Although simulated canals guarantee the

standardization, but it is different from natural teeth in dentin hardness and canal irregularities(11)(12)(13). Also, the heat produced by rotary instruments through friction may cause resin to melt(14)(15). Mesio-buccal canals of mandibular molars were selected because they show curvature at least in mesiodistal plane. They are narrow increasing instrumentation difficulty. (16)

Various methods are used to evaluate the quality of endodontic instrumentation including, scanning electron microscope, photography, a prefabricated model of the canal and computerized systems(17). Computed tomography (CT) and micro-CT were first suggested for this purpose since they are non-invasive and do not require tooth sectioning. CBCT has been confirmed to be an accurate tool to study the root canal anatomy(18)(19).

In the present study all groups showed significant reduction in canal curvature. The highest value of percentage of reduction was found with samples shaped with Revo-S followed by 2Shape while the lowest value were found in samples shaped with M-pro.Revo-S is made of conventional Ni-Ti alloy increasing file tendency to straightening canal curve .While M-pro and 2Shape are made of heat treated Ni-Ti alloy. Since heat-treatment increases file flexibility allowing better respect to original canal curve. (20)(21) (22) .

**Conclusion**

With the limitation of the present study it was concluded that Revo -S root canal rotary system caused the highest percent of reduction in canal curve among all groups.

## References

1. Schäfer E, Schulz-Bongert U, Tulus G. Comparison of hand stainless steel and nickel titanium rotary instrumentation: A clinical study. *J Endod.* 2004 Jun 1;30(6):432–5.
2. González-Rodríguez MP, Ferrer-Luque CM. A comparison of Profile, Hero 642, and K3 instrumentation systems in teeth using digital imaging analysis. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2004 Jan 1;97(1):112–5.
3. Micro Mega. 2shape [Internet]. [cited 2020 Jan 5]. Available from: <http://www.denta.be/pdf/MicroMega-2shape-ENG.pdf>
4. innovative materials and devices [Internet]. [cited 2020 Jan 5]. Available from: <http://www.imdmedical.com/#/Products/11>
5. Schneider SW. A comparison of canal preparations in straight and curved root canals. *Oral Surgery, Oral Med Oral Pathol.* 1971;32(2):271–5.
6. Mcspadden JT. Mastering endodontic instrumentation. Chattanooga, TN: Cloudland Institute. 2007. 21–28 p.
7. Aguiar CM, Sobrinho PB, Teles F, Câmara AC, de Figueiredo JAP. Comparison of the centring ability of the ProTaper and ProTaper Universal rotary systems for preparing curved root canals. *Aust Endod J.* 2013;39(1):25–30.
8. Weine FS, Kelly RF, Lio PJ. The effect of preparation procedures on original canal shape and on apical foramen shape. *J Endod.* 1975;1(8):255–62.
9. Glosson CR, Haller RH, Brent Dove S, et al. A comparison of root canal preparations using Ni-Ti hand, Ni-Ti engine-driven, and K-Flex endodontic instruments. *J Endod.* 1995;21(3):146–51.
10. Hulsmann M, Peters OA, Dummer PMH. Mechanical preparation of root canals: shaping goals, techniques and means. *Endod Top.* 2005;10(1):30–76.
11. Lim Kc, Webber J. The validity of simulated root canals for the investigation of the prepared root canal shape. *Int Endod J.* 1985;18(4):240–6.
12. Spyropoulos S, Eldeeb Me, Messer HH. The effect of Giromatic files on the preparation shape of severely curved canals. *Int Endod J.* 1987;20(3):133–42.
13. Miserendino LJ, Miserendino CA, Moser JB, et al. Cutting efficiency of endodontic instruments. Part III. Comparison of sonic and ultrasonic instrument systems. *J Endod.* 1988;14(1):24–30.
14. Deepak J, Ashish M, Patil N, et al. Shaping Ability of 5(th) Generation Ni-Ti Rotary Systems for Root Canal Preparation in Curved Root Canals using CBCT: An In Vitro Study. *J Int oral Heal.* 2015;7(Suppl 1):57–61.
15. Kum KY, Spängberg L, Cha BY, et al. Shaping ability of three profile rotary instrumentation techniques in simulated resin root canals. *J Endod.* 2000;26(12):719–23.
16. Singh S, Gupta T, Pandey V, et al. Shaping ability of two-shape and ProTaper Gold Files by using Cone-beam computed tomography. *J Contemp Dent Pract.* 2019;20(3):330–4.
17. Bramante CM, Berbert A, Borges RP. A methodology for evaluation of root canal instrumentation. *J Endod.* 1987;13(5):243–5.
18. James M. Gambill, Marden Alder CE del R. Comparison of nickel-titanium and stainless steel hand-file instrumentation using computed tomography. *J Endod.* 1996;22(7):369–75.
19. Peters OA, Laib A, Göhring TN, et al. Changes in root canal geometry after preparation assessed by high-resolution computed tomography. *J Endod.* 2001;27(1):1–6.
20. Saber SEDM, Nagy MM, Schäfer E. Comparative evaluation of the shaping ability of WaveOne, Reciproc and OneShape single-file systems in severely curved root canals of extracted teeth. *Int Endod J.* 2015;48(1):109–14.
21. Alapati SB, Brantley WA, Iijima M, et al. Metallurgical Characterization of a New Nickel-Titanium Wire for Rotary Endodontic Instruments. *J Endod.*

- 2009;35(11):1589–93.
22. **Lopes HP, Gambarra-Soares T, Elias CN, et al.** Comparison of the mechanical properties of rotary instruments made of conventional nickel-titanium wire, M-Wire, or nickel-titanium alloy in R-phase. *J Endod.* 2013 Apr 1;39(4):516–20.



